

Efficacy of Commonly Used 3D Mapping Systems in Acute Success Rates of Catheter Ablation Procedures

George Bazoukis,^{1,2} Khaled Elkholey,³ Stavros Stavrakis,³ E Kevin Heist⁴ and Antonis A Armoundas^{5,6}

1. School of Medicine, European University Cyprus, Nicosia, Cyprus; 2. Department of Cardiology, Larnaca General Hospital, Larnaca, Cyprus; 3. Heart Rhythm Institute, University of Oklahoma Health Sciences Center, Oklahoma City, OK, USA; 4. Cardiology Division, Cardiac Arrhythmia Service, Massachusetts General Hospital, Boston, MA, USA; 5. Cardiovascular Research Center, Massachusetts General Hospital, Boston, MA, USA; 6. Broad Institute, Harvard University and Massachusetts Institute of Technology, Cambridge, MA, USA

Introduction: This systematic review aims to summarize the procedural arrhythmia termination rates in catheter ablation (CA) procedures of atrial or ventricular arrhythmias using the commonly used mapping systems (CARTO, Rhythmia and EnSite/NavX). **Materials and Methods:** A systematic search in MEDLINE and Cochrane databases through February 2021 was performed. **Results:** With regard to atrial fibrillation ablation procedures, acute success rates ranged from 15.4 to 96.0% and 9.1 to 100.0% using the CARTO and EnSite/NavX mapping systems, respectively; acute atrial tachycardia (AT) termination to sinus rhythm ranged from 75 to 100% using the CARTO system. The acute success rate for different types of AT ranged from 75 to 97% using Rhythmia, while the NavX mapping system was also found to have excellent efficacy in the setting of AT, with acute arrhythmia termination rates ranging from 73 to 99%. With regard to ventricular tachycardia, in the setting of ischaemic cardiomyopathy, acute success rates ranged from 70 to 100% using CARTO and 64% using EnSite/NavX systems. The acute success rate using the Rhythmia system ranged from 61.5 to 100.0% for different clinical settings. **Conclusions:** Mapping systems have played a crucial role in high-density mapping and the observed high procedural success rates of atrial and ventricular CA procedures. More data are needed for the comparative efficacy of mapping systems in acute arrhythmia termination, across different clinical settings.

Keywords

Atrial fibrillation, CARTO, catheter ablation, EnSite/NavX, mapping systems, Rhythmia, ventricular tachycardia

Disclosures: E Kevin Heist reports the following relationships with industry: Abbott (consulting, honorarium, research grants), Biotronik (honorarium, research grants), Boston Scientific (consulting, honorarium, research grants), Medtronic (consulting, honorarium), Oracle Health (equity), Pfizer (consulting). George Bazoukis, Khaled Elkholey and Stavros Stavrakis have no financial or non-financial relationships or activities to declare in relation to this article. Antonis Armoundas has received research funding support as detailed below.

Review process: Double-blind peer review.

Compliance with ethics: This article involves a review of the literature and does not involve any studies with human or animal subjects performed by any of the authors.

Data availability: Data sharing is not applicable to this article as no datasets were generated or analyzed during the writing of this article.

Authorship: The named authors meet the International Committee of Medical Journal Editors (ICMJE) criteria for authorship of this manuscript, take responsibility for the integrity of the work as a whole and have given the final approval for the version to be published.

Access: This article is freely accessible at touchCARDIO.com © Touch Medical Media 2024.

Received: 23 November 2023

Accepted: 12 December 2023

Citation: *Heart International*. 2024;18(1):Online ahead of journal publication

Corresponding author: Antonis A Armoundas, Cardiovascular Research Center, Massachusetts General Hospital, 149 13th Street, Charlestown, MA 02129, USA. E: Armoundas.Antonis@mgh.harvard.edu

Support: Antonis A Armoundas has been supported by the RICBAC Foundation, NIH grants 1 R01 HL135335-01, 1 R21 HL137870-01, 1 R21EB026164-01 and 3R21EB026164-02S1. The journal and journal publisher have not received any support for the article.

Catheter ablation (CA) is an invasive adjunctive therapeutic option for atrial and ventricular arrhythmias, in uncontrolled cases, following optimal medical therapy.^{1,2} The validation of 3D mapping and its implementation into clinical usage has been a remarkable achievement in the field of complex ablation of scar-related ventricular tachycardia (VT) in the early 2000s.³ In the setting of atrial fibrillation (AF), 3D navigation systems have been crucial for reducing complications, including pulmonary vein (PV) stenosis, perforation, phrenic nerve or oesophageal injury, during pulmonary vein isolation (PVI), or substrate modification, defragmentation or linear ablations, during chronic AF ablation.⁴⁻⁶

While clinical characteristics and the operators' experience can affect long-term success rates of CA, the acute success rate is less dependent on clinical variables during the follow-up period, and therefore, it is a better marker of the efficacy of the different mapping systems.⁷⁻⁹ This systematic review aims to summarize the existing data about the acute arrhythmia termination success rates in atrial and ventricular CA procedures using the most common mapping systems (CARTO [Biosense Webster Inc., Diamond Bar, CA, USA], Rhythmia [Boston Scientific, Natick, MA, USA], CardiInsight [Medtronic Inc, Minneapolis, USA] and NavX/EnSite [St. Jude Medical, St Paul, Minnesota, USA]).

Materials and methods

This systematic review was performed in accordance with the Preferred Reporting Items for Systematic Reviews and Meta-Analyses statement.¹⁰

Search strategy

This study aimed to identify all relevant studies that provided data about the acute success rates of different mapping systems in patients who underwent the CA procedure either for AF and atrial tachycardia (AT) or for VT and premature ventricular contractions (PVCs). Two independent investigators performed a systematic search in MEDLINE and Cochrane

databases through July 2021. The reference lists of the included studies, as well as relevant review studies, were manually searched. The following keywords were used to retrieve all relevant studies: "(CARTO OR Rhythmia OR NavX OR CardiInsight OR EnSite) AND (atrial OR ventricular) AND (fibrillation OR tachycardia OR flutter)", without any limitations. We first screened the titles and abstracts of each retrieved study, and if a study was considered relevant, then the full text was studied.

Inclusion/exclusion criteria

We included studies that provided data on acute success rates during ablation procedures in either AF/AT or VT/PVCs across the different mapping systems. Acute success is defined either as clinical arrhythmia termination or as any arrhythmia termination or non-inducibility. The definitions of arrhythmia termination that were used in each study are presented in *Tables 1 and 2*.

We excluded studies that did not report data about the acute success rates or the type of mapping system that was used for the ablation procedure. Furthermore, we excluded studies including only patients with congenital heart diseases, heart transplant recipients and children. Case reports/series, animal studies, studies on ventricular fibrillation ablation procedures and studies on hybrid ablation procedures have also been excluded.

Data extraction

The data extraction was performed independently by two authors (GB and KE). The following data were extracted: first author, year of publication, journal name, type of study (single or multicentre), type of arrhythmia (AF, AT, VT or PVCs), duration of follow-up, number of patients, gender, age, type of cardiomyopathy, left ventricular ejection fraction, type of mapping system, type of catheters, mapping techniques, acute success definition, acute procedural success rates, procedural complications, follow-up, arrhythmia recurrence during follow-up, fluoroscopy time and type of ablation (endocardial, epicardial or combined).

The quality assessment of the included studies was performed using the NIH Quality Assessment Tool.⁸⁰

Results

Study search

Of the 848 studies, 661 were excluded at the title/abstract level and 117 were excluded at the full-text level. Finally, 70 unique studies were included in the systematic review. Of them, 42 studies reported data about the acute procedural success of atrial CA procedures, while 29 studies provided data about the acute procedural success of ventricular CA procedures. One study provided data on both ventricular and atrial CA procedures.¹⁵ The search strategy is shown in Figure 1.

Study characteristics

The baseline characteristics and the main findings of the included studies are presented in *Tables 1 and 2* and in Figure 2. With regard to atrial arrhythmias, the search strategy revealed 17 unique studies that used the CARTO mapping system.^{11–26} Specifically, 10 studies provided data about AF ablation, 1 study on atrial flutter (AFL) ablation and 8 studies on AT ablation (one study provided data on all three types of arrhythmias).^{11–26} The Rhythmia mapping system was used in seven studies on atrial arrhythmias (two AF studies and five AT studies).^{27–33} Finally, in the setting of atrial arrhythmias, 20 studies provided data on the NavX/EnSite mapping systems (16 AF studies and 4 AT studies – 2 studies also provided data on the CARTO mapping system and therefore were included in both systems).^{12,20,34–51} As each of AF, AFL and AT has

distinct electrophysiological mechanisms and substrates, we elected to report outcomes for each of these arrhythmias separately.

In the setting of ventricular CA, the search strategy revealed 20 studies that used the CARTO mapping system (12 studies in patients with ischaemic cardiomyopathy, 4 studies on outflow tract VT, 2 studies in the setting of arrhythmogenic right ventricular cardiomyopathy [ARVC] and 2 studies in mixed population); 4 studies that used the Rhythmia mapping system in mixed population including ischaemic and non-ischaemic cardiomyopathy, hypertrophic cardiomyopathy and congenital heart diseases and 5 studies that used the NavX/EnSite and CardiInsight mapping systems (one study in patients with ischaemic cardiomyopathy, one study in patients with ARVC, one study in patients with outflow tract VT and two studies in mixed population).^{15,52–69,71–79}

All studies achieved a good quality rating according to the National Heart, Lung and Blood Institute study quality assessment tool (*Online Supplement Table S1*).⁸¹

Outcome data

Atrial arrhythmias

CARTO mapping system

In the clinical setting of persistent AF, CARTO exhibited a variable performance (10 studies) in terms of acute arrhythmia termination (*Table 1*). Specifically, in this setting, the acute success rates ranged from 15.4 to 96%, depending mainly on the type of AF and the definition of acute success (termination to sinus rhythm [SR] or termination of AF to any other organized rhythms).^{11–14,16–19,82}

Furthermore, in the included studies, CARTO was used for the management of AT (eight studies). In this setting, the acute termination of ATs to SR ranged between 75 and 100%.^{20–26} Similarly, one study showed that the acute termination rate for AFL was 95% (*Table 1*).¹⁵

Rhythmia mapping system

Our search identified seven studies that used the Rhythmia mapping system for guiding CA procedures in patients with AT. These data showed excellent efficacy in terms of acute arrhythmia termination rates (*Table 1*).^{27–33}

Furthermore, the search strategy revealed that the acute success rate for different types of AT (seven studies) ranged from 75 to 97%.^{28–33}

NavX/EnSite mapping systems

In the setting of AF ablation, the use of the NavX/EnSite mapping system resulted in acute arrhythmia termination rates (16 studies), ranging between 9.1 and 100%, depending mainly on the type of AF (*Table 1*).^{12,34–48} However, it should be mentioned that the study that showed the lowest arrhythmia termination rate provided data about arrhythmia termination with PVI only in patients with persistent AF.⁴⁶

The NavX mapping system was also found to have excellent efficacy in the setting of AT (three studies), with acute arrhythmia termination rates ranging between 73 and 99% (*Table 1*).^{49–51,83}

Ventricular arrhythmias

CARTO mapping system

The CARTO mapping system has been used for the management of ventricular arrhythmias (20 studies) in different clinical settings (*Table 2*).^{15,52–70} The identified studies showed that the CARTO mapping system had an excellent efficacy in ischaemic cardiomyopathy (12

Table 1: Baseline characteristics and main outcomes of the included studies related to atrial arrhythmia catheter ablation procedures

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the 1/1 (%)	Fluoroscopy time (min) mean ± SD
Atrial arrhythmias																
CARTO mapping system																
Calvo et al. (2017) ¹¹	Spain	Single	13	Pers AF (long standing)	53.9	100	57.7	CARTO	2 (15.4)	SR restoration	Phase/frequency mapping	Irrigated Navistar ThermoCool SmartTouch, Penatray and a quadripolar catheter (Biosense Webster)	No complications	12-72	30	17.7±3.9
Choo et al. 2011 ¹²	UK	Single	47	PAF and Pers AF	56.2	68	58.9	CARTO	45 (96)	SR restoration	-	Irrigated (Celsius or Navistar ThermoCool [Biosense Webster] or Cool Path Duo [St. Jude Medical, St. Paul, MN, USA])	Cardiac tamponade (n=1)	12	60	73±27
Nademanee et al. (2004) ¹³	Thailand	Single	121	PAF and Perm AF	63	76	n/a	CARTO	115 (95)	Complete elimination of the areas with CFAEs or conversion of AF into normal SR for patients with both CAF and PAF	Voltage maps	Standard 4 mm tip catheter	Cerebrovascular accident (n=1), cardiac tamponade (n=2), complete atrioventricular block (n=1), transient severe pulmonary oedema (n=1) and femoral arterial atrioventricular fistula (n=1)	12	9	14.7±4.8
Seitz et al. (2011) ¹⁴	EU	Single	22	NPAF: 86.4%	56.6	82	44.7	CARTO (Biosense Webster, Diamond Bar, CA, USA)	21 (95.5%)	AF termination defined as the conversion of AF into SR or regularization into stable AT	CFAE LA map	2-5-2 mm electrode spacing (Xtrem; ELA Medical, Le Plessis-Robinson, France), 3.5 mm irrigated-tip quadripolar ablation catheter (2-5-2 mm inter-electrode spacing, ThermoCool; Biosense Webster)	None	12	18	39±15
Seitz et al. (2013) ¹⁴	EU	Single	32	NPAF: 90.6%	60.4	75	52.6	CARTO (Biosense Webster)	30 (93.7)	AF termination defined as the conversion of AF into SR or regularization into stable AT	CFAE LA map	Decapolar catheter (2-5-2 mm electrode spacing, Xtrem; ELA Medical), 3.5 mm irrigated-tip quadripolar ablation catheter (2-5-2 mm inter-electrode spacing, ThermoCool [Biosense Webster] F or J curve)	None	12	15.6	31±17
Suleiman et al. (2007) ¹⁵	Israel	Single	13	AF	49 [†]	59 [†]	n/a	CARTO system (Biosense Webster, Johnson & Johnson, USA)	12 (92)	Termination to SR	-	-	Transient cerebrovascular accident (n=1)	33* refers to the total population	38.5	10.3±6.9* refers to the total population
Takahashi et al. (2018) ¹⁶	Japan	Single	68	AF	62	68	59	CARTO 3 (Biosense Webster)	35 (51.5)	Termination into SR	Activation mapping	Navistar ThermoCool SF catheter (Biosense Webster)	None	17.3	42.7	n/a

Continued

Table 1: Continued

Study (year)	Country	Single/multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the 1/10 (%)	Fluoroscopy time (min) mean ± SD
Weng et al. (2009) ⁷	China	Single	111	Pers AF	50.8	55.9	53.1	CARTO (Biosense Webster)	SR → 30 (27) AT → 35 (31.5)	Termination to SR or AT (both data are provided)	Activation mapping and CFAE mapping	Decapolar catheter (Biosense Webster), mapping catheter (Lasso; Biosense Webster) and saline-irrigated 3.5 mm catheter (ThermoCool Navistar; Biosense Webster)	Catheter entrapment in valve prosthesis without further complications (n=1) and major stroke (n=1)	4.1 med	46.9	24±15
Wnuk-Wojnar et al. (2005) ⁸	Poland	Single	94	PAF: 63.8%	54.3	64.9	58	CARTO (Johnson & Webster)	65 (69.6)	Termination to SR	Disappearance of electrical potentials within ablation lines and consequently inability to induce and maintain AF	n/a	Stroke (n=1), PV thrombosis (n=1), HIT tamponade (n=2)	12 med	30	22.4 (11–41)
Wu et al. (2008) ⁹	EU	Single	10	Pers AF	62.3	50	n/a	CARTOX (Biosense Webster)	3 (30)	Termination of AF and/or prolongation of AFL	CFAE mapping	4-polar distally circular catheter (OrbiterPV; C.R. Bard), octapolar reference catheter (XPTM; C.R. Bard Electrophysiology, Lowell, MA, USA), 4 mm irrigated tip catheter (Navistar ThermoCool; Biosense Webster)	n/a	n/a	n/a	
Suleiman et al. (2007) ⁵¹	Israel	Single	37	AFL	49 [†]	59 [†]	n/a	CARTO system (Biosense Webster, Johnson & Johnson)	35 (95)	Termination to SR	-	-	None	33* refers to the total population	10.8	10.3±6.9* refers to the total population
Esata et al. (2009) ⁵²	EU	Single	26 (12 CARTO, 14 NavX)	Macro-reentrant AT	59	76.9	56	CARTO (Biosense Webster, Inc.) or NavX (Endocardial Solutions, Inc., St. Paul, MN, USA)	26 (100%) for both systems	Tachycardia termination and non-inducibility	Entrainment	Irrigated, F-type irrigated-tip, Navistar ThermoCool (Biosense Webster); W-Curve IBI Therapy Cooled Path (St. Jude Medical)	No complications	10.1	12	37±19
Jamil-Copley et al. (2013) ²¹	UK	Single	10	AT	57.7	70	n/a	CARTO-XP system (Biosense Webster, Haifa, Israel)	9 (90)	Increase in TCL	Ripple, activation and entrainment	-	n/a	n/a	n/a	n/a
Okumura et al. (2016) ²²	Japan	Single	6	AT	67	33.3	n/a	CARTO 3 system (Biosense Webster)	6 (100)	Termination of AT and no inducibility of AT after isoproterenol infusion	Activation and entrainment	7F 4 mm tip (Navistar; Biosense Webster), SF quadripolar catheter (5–5 mm, inquiry catheter; St. Jude Medical), SF decapolar catheter (Fe-po; Fukuda Denshi Co., Ltd., Tokyo, Japan) and 7F steerable duodecapolar Halo catheter (inquiry catheter; St. Jude Medical)	None	11	0	n/a
Suleiman et al. (2007) ⁵⁴	Israel	Single	20	AT	49 [†]	59 [†]	n/a	CARTO system (Biosense Webster, Johnson & Johnson)	15/20 (75)	Termination to SR	-	-	None	33* refers to the total population	25	10.3±6.9* refers to the total population

Continued

Table 1: Continued

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the 1/1 (%)	Fluoroscopy time (min) mean \pm SD	
Strisciuglio et al. (2019) ²⁸	EU	Single	31	AT	69	63	n/a	CARTO	SR \rightarrow 23 (66) AT \rightarrow 12 (34)	AT termination to SR or AT with different cycle lengths	High-density activation mapping and entrainment mapping	Multi-electrode mapping catheter (Pentaray, Biosense Webster Inc, Irvine, CA, USA) and an open-tip irrigated RF catheter (8F) with tip-integrated contact force sensor (ThermoCool SmartTouch, Biosense Webster Inc)	n/a	n/a	n/a	n/a	
Vicera et al. (2020) ⁴	Taiwan	Single	20	AT	58.2	75	59.6	CARTO 3 system CONFIDENSE Module	SR \rightarrow 18 (69.2) Another AT \rightarrow 7 (24.9)	Termination or change to a different activation pattern with an associated change in cycle length and negative inducibility of clinical AT	Activation mapping and coherent map	Multielectrode mapping catheter (Lasso Nav or Pentaray Nav, Biosense Webster, Inc.)	n/a	13.7	25	n/a	
Wo et al. (2014) ²⁵	Taiwan	Single	15	AT	49	66.7	62	CARTO (Biosense Webster)	15 (100)	Termination of MRAT and non-inducibility	Entrainment mapping	4 mm Nav/Star catheter (irrigating or non-irrigating; Biosense Webster)	n/a	15	0	n/a	
Yagshita et al. (2019) ²⁶	Japan	Single	39	AT	65	78	59	CARTO 3 CONFIDENSE Module (Biosense Webster)	90	Termination to SR	LA1, voltage and entrainment	Mapping catheter (Pentaray Nav, Biosense Webster), 3.5 mm irrigation catheter (ThermoCool; Biosense Webster)	n/a	n/a	n/a	n/a	
Rhythmia mapping system																	
Kiamura et al. (2018) ⁷	EU	Two-centre	8	Bilatral AT	59.5	87.5	59.9	Rhythmia (Boston Scientific, Marlborough, MA, USA)	8 (88.9)	Termination of biatral AT to SR or to another AT	High-density activation map	3.5 mm tip open-irrigated catheter (ThermoCool SF, Biosense Webster), decapolar catheter within the coronary sinus (Extreme, Sorin/Dynamic, XT, Boston Scientific)	n/a	12.4 med	12.5	n/a	n/a
Yamashita et al. (2019) ²⁹	EU and Japan	Multicentre	26	Post-AF ablation AT	63	69	57	Rhythmia (Boston Scientific)	25 (96.2)	Restoration of SR or a change in the AT by RF ablation of the PV gap	Activation and entrainment	64-pole mini-basket catheter (Orion; Boston Scientific), non-magnetic open-irrigated catheter (Celsius, ThermoCool; Biosense Webster, Inc.)	n/a	12	15	n/a	n/a
Anter et al. (2016) ²⁹	US	Multicentre	20	Recurrent AT	62	n/a	n/a	Rhythmia (Boston Scientific, Cambridge, MA, USA)	18 (75)	Termination to SR	Activation, entrainment and pace	Orion basket catheter (Boston Scientific), nonmagnetic open-irrigated catheter (Celsius; ThermoCool; Biosense Webster)	None	7.5	25	n/a	n/a
De Simone et al. (2020) ³⁰	EU	Single	24	Re-entrant AT	54	54.2	n/a	Rhythmia (Boston Scientific, Inc.)	96.2 [†]	Termination to SR	Propagation map and entrainment	64-pole mini-basket mapping catheter (IntellaMap Orion; Boston Scientific, Inc.)	None	18	12.5	n/a	n/a

Continued

Table 1: Continued

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the f/u (%)	Fluoroscopy time (min) mean ± SD
Latcu et al. (2017) ³¹	EU	Single	19	AT	71 med	68.4	62 med	Rhythmia (Boston Scientific)	29 (97)	Termination into SR or another stable AT	Activation maps	Decapolar diagnostic catheter (Inquiry L, 2–5–2 mm spacing; Saint Jude Medical), F-type, 2–8–2 mm spacing (Biosense Webster) and an IntelMap Orion mapping catheter	None	12 med	16	18±10
Takigawa et al. (2017) ³²	EU	Single	57	AT	61.9	82.5	54.3	Rhythmia (Boston Scientific, Natick, MA, USA)	66/88 (75) [†]	AT termination	Activation and entrainment	Orion multipolar basket catheter (Boston Scientific), 3.5 mm tip ablation catheter (ThermoCool SF catheter; Biosense Webster)	n/a	6	26–3	n/a
Takigawa et al. (2018) ³³	EU	Single	41	Post-AF ablation AT	65.7	70.7	52.6	Rhythmia system (Boston Scientific)	80.6 [†]	AT termination	Activation	Orion multipolar basket catheter; 3.5 mm tip ablation catheter (ThermoCool SF; Biosense Webster)	n/a	12	45.2	n/a
NavX-EnSite mapping systems																
Aksu et al. (2020) ³⁴	Turkey and USA	Single	12	Pers AF	53	83	48.5	EnSite Precision (Abbott Medical, Chicago, IL, USA)	8 (66.7)	SR restoration	Fractionation and mapping	3.5 mm tip ablation catheter (TractiCath™ or FlexAblity™; Abbott), 20-pole spiral double-loop catheter (Inquiry™, AFocus II™ (Abbott), 1 mm length, 4 mm spacing, 20 mm fixed-loop diameter)	n/a	34 med	25	n/a
Miyamoto et al. (2009) ³⁵	Japan	Single	50	Pers AF; 20%	61.2	74	n/a	EnSite version 6.0J (St. Jude Medical)	75.6	AF termination to SR during PVAI alone or additional RF ablation	Voltage and activation maps	8 mm tip ablation catheter	n/a	14	8%	n/a
Miyamoto et al. (2010) ³⁶	Japan	Single	20	Pers AF; 20%	58	70	n/a	EnSite version 6.0J (St. Jude Medical)	14 (70)	AF termination to SR	CFAE ablation	7F catheter with an 8 mm tip distal electrode (Fantasia, Japan Lifeline Co. Ltd., Tokyo, Japan)	n/a	n/a	n/a	n/a
Yamaguchi et al. (2010) ³⁷	Japan	Single	65	PAF	58	84.6	n/a	EnSite array	47/51 (92.2)	AF termination to SR	Activation map	20-pole circular mapping catheter (Optima; St. Jude Medical) and deflectable 7F quadripolar, non-irrigated 8 mm tip electrode ablation catheter (F-anastasia; Japan Lifeline)	No complications	23	16.9	n/a
Ammar-Busch et al. (2018) ³⁸	EU	Single	16	Pers AF	63	88	n/a	NavX (St. Jude Medical), ECOVUE™ (Cardionight Technologies, Inc., Cleveland, OH, USA)	12 (75)	AF termination	CFAE mapping	14 bipoles Orbiter PV, C.R. Bard or 20 bipoles AFocus II™ (St. Jude Medical), irrigated tip catheter (Therapy™ Cool Flex™; St. Jude Medical)	No complications	n/a	n/a	n/a

Continued

Table 1: Continued

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the 1/10 (%)	Fluoroscopy time (min) mean ± SD
Kumagai et al. (2013) ³⁹	Japan	Single	50	PAF and Pers AF	63.9	78	61	NavX system (NavX with CFE software; St. Jude Medical Inc.)	18 (36)	Termination to SR	CFAE, activation and mapping	7 F decapolar circular catheter (Lasso; Biosense Webster, Inc) and 3.5 mm irrigated tip RF catheter (Safire; St. Jude Medical Inc.)	n/a	12	28	n/a
Kumagai et al. (2017) ⁴⁰	Japan	Single	32	AF	57	87.5	63	NavX system (NavX with CFE software; St. Jude Medical Inc.)	26 (81.3)	AF termination or >10% slowing of the AF cycle length from after the PVI to the end of the high-DI and continuous CFAE- site ablation	CFAE, activation and mapping	7 F decapolar circular catheters (Lasso; Biosense Webster, Inc.), 3.5 mm irrigated tip radiofrequency catheter (Safire; St. Jude Medical Inc.) and a 20-pole mapping circular catheter (St. Jude Medical Inc.)	No complications	12	19	n/a
Choo et al. (2011) ¹²	UK	Single	24	PAF and Pers AF	62.2	71	55.4	NavX	24 (100)	SR restoration	-	Irrigated (Celsius or NavStar ThermoCool (Biosense Webster) or Cool Path Duo (St. Jude Medical))	Pulmonary vein puncture (n=1)	-	-	79±25
Lin et al. (2009) ⁴¹	Japan	Single	60	NPAF	49	83.3	55	NavX (with CFE software; St. Jude Medical Inc.)	20/60 (33.3)	AF restored to SR during ablation	CFE mapping	Irrigated-tip 4 mm ablation catheter (EPT; Boston Scientific Corporation), circular catheter recording (Spiral, AF Division; St. Jude Medical Inc.)	Patient with cardiac tamponade (n=1)	19	25	n/a
Lo et al. (2009) ⁴²	Japan	Single	87	Perm AF	53	82.8	54	NavX (St. Jude Medical)	30 (34.5)	Termination to SR	CFAE mapping	Irrigated-tip 3.5 mm ablation catheter (Chill i™, EPT; Boston Scientific Corporation, San Jose, CA, USA) and catheter recording (Spiral, AF Division; St. Jude Medical)	Pericardial effusion needing pericardiocentesis (n=1)	21	21	n/a
Matsuo et al. (2012) ⁴³	Japan	Single	40	Pers AF	53.5	97.5	62.4	EnSite NavX (St. Jude Medical)	13 (32.5)	Termination to SR	CFAE mapping	16-polar tow site (6-polar for the right atrium and 10-polar mapping catheter (Inquiry Luna-Cath; St. Jude Medical)), 20-polar, circumferential mapping catheter of 20, 25 or 30 mm in diameter (Inquiry Optima; St. Jude Medical or Lasso, Biosense Webster) and irrigated 3.5 mm tip ablation catheter (CoolPath™ Duo (St. Jude Medical) or Thermocool Navistar (Biosense Webster))	n/a	19.7	20	n/a
Nair et al. (2009) ⁴⁴	India	Single	21	Perm AF	44	42.9	n/a	NavX EnSite software, version 7, at St. Jude Medical	18 (85.7)	Organization of the atrial electrograms or conversion into an SR	CFAE mapping and voltage maps	2 mm irrigated tip ablation catheter (IBI Therapy Cool Path Ablation Catheter; Irvine Biomedicals, Irvine, CA)	No complications	9.8	23.1	n/a

Continued

Table 1: Continued

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the f/u (%)	Fluoroscopy time (min) mean \pm SD
Nakahara et al. (2014) ⁴⁵	Japan	Single	60	Pers AF	63.1	83	59	NavX, with CFE software (St. Jude Medical, Inc.)	19 (31.7)	Termination of AF to SR directly or via one or more intermediate ATs	CFAE mapping	20-pole, 15–25 mm Lasso catheter with 6 mm bipole spacing (Biosense Webster), 4 mm irrigated-tip catheter (Safire; St. Jude Medical) and 20-pole, 20 mm AFocus II catheter with 4 mm bipole spacing (St. Jude Medical)	n/a	16 med	21.7	44.8 \pm 8.1
Roux et al. (2009) ⁴⁶	USA	Single	22	Pers AF	58	82	60	NavX system (NavX, St. Jude Medical)	2 (9.1)	AF termination with PVI	CFAE mapping	Mapping catheter (10- pole, 15–25 mm Lasso, 6 mm bipole spacing; Biosense Webster) and an 8 or 3.5 mm irrigated-tip catheter (Biosense Webster)	n/a	n/a	n/a	99 \pm 35
Suenari et al. (2011) ⁴⁷	Taiwan	Single	23	PAF	54.3	83	59	EnSite NavX (St. Jude Medical, Inc.)	16/23 (69.6)	Termination of AF to SR directly or through one or more intermediate ATs	CFAE mapping	Irrigated 3.5 mm tip ablation catheter (EP1; Boston Scientific Corporation), 7F 10-pole/6F 4-pole catheters (St. Jude Medical Inc) and 5F 12- pole circular catheter (Spiral AF Division; St. Jude Medical Inc.)	n/a	11.9	24	n/a
Verma et al. (2011) ⁴⁸	US	Single	30	Pers AF	63	75	52	EnSite NavX mapping system (St. Jude Medical)	4 (13.3) 3 (10%) (termination to SR)	AF termination to SR or AT	CFAE mapping	Circular mapping catheter with ten 2 mm electrodes and 2 mm interelectrode spacing (Lasso; Biosense Webster) and 3.5 mm tip ablation catheter (ThermoCool; Biosense Webster)	Left femoral haematoma (n=1)	12	43	55 \pm 20
Narita et al. (2010) ⁴⁹	Japan	Single	51	AT	57	54.9	n/a	EnSite Multielectrode array (version 3.2 in 33 patients; version 6.0 in 18 patients)	99	AT termination to SR and subsequent non-inducibility	Voltage and activation maps	A 4 or an 8 mm tip steerable catheter (Famadis Japan LineIn; Blazer II (Boston Scientific))	No complications	16	3.9	19 \pm 11

Continued

Table 1: Continued

Study (year)	Country	Single/ multicentre	N	Type of arrhythmia	Mean age (years)	Males (%)	EF (%)	Mapping system	Acute success, N (%)	Acute success definition	Mapping technique	Catheter type	Complications	Follow-up (months)	Recurrence at the 1/y (%)	Fluoroscopy time (min) mean ± SD
Patel et al. (2009) ³⁹	USA	Single	17	Post-AF ablation/AT	62	82.4	61	EnSite NavX	96 [†]	AT termination to SR and non- inducibility	Activation and entrainment	3.5 mm-tip Celsius ThermoCool (Biosense Webster, Inc.) or Chili (Boston Scientific, Inc.), mapping catheter (Lasso or Optima catheters; Biosense Webster, Inc., and St Jude Medical, Inc.), PentaRay catheter is a 20-pole steerable mapping catheter arranged in five soft radiating spines (1 mm electrodes separated by 4, 4, and 4 mm inter-electrode spacing) covering a diameter of 3.5 cm (PentaRay; Biosense Webster, Inc.) and irrigated ablation catheter	No complications	7	23.5	n/a
Nagamoto et al. (2011) ²¹	Japan	Single	33	Post-AF ablation/AT	59	n/a	n/a	EnSite version 6.0 (St. Jude Medical)	24 (73)	Termination of AT or change to another AT or non- inducibility	CFAE ablation	Livewire (St. Jude Medical), multielectrode array catheter (St Jude Medical), 20-pole circular electrode catheter (Optima, St. Jude Medical) and non-irrigated ablation catheter with an 8 mm tip (Fanzasia, Japan Lifeline)	Femoral arteriovenous fistula (n=1)	21	9	n/a

[†]Choo's study provided separate data about the patient characteristics for CARTO and NavX mapping systems.

[‡]Suleiman's study provided data on AF, AFL and AT, while the baseline characteristics are provided only for the total population.

[§]The denominators in these percentages are the total number of atrial arrhythmias and not the total number of the included patients.

[¶]These are similar cohorts, and therefore, these data were used once in the quantitative synthesis.

AF = atrial fibrillation; AFL = atrial flutter; AT = atrial tachycardia; CFAE = complex fractionated atrial electrogram; EF = left ventricular ejection fraction; F = French; 1/y = follow-up; Med = median; n/a = not available; NPAF = non-paroxysmal AF; PAF = paroxysmal AF; Perm AF = permanent AF; Pers AF = persistent AF; PV = pulmonary vein; PVI = pulmonary vein isolation; PVI = pulmonary vein isolation; SR = sinus rhythm.

Table 2: Baseline characteristics and main outcomes of the included studies related to ventricular arrhythmia catheter ablation procedures

Study (year)	Country	Single/multicentre	N	Mean age (years)	Males (%)	Clinical setting	EF (%)	Mapping system	Mapping type	Acute success/definition of success	Complications (n)	Catheter	Follow up (months)	Recurrence at follow up (%)	Endo or epi or both	Fluoroscopy time (min) mean±SD	
Ventricular tachycardia – PVCs																	
CARTO mapping system																	
Luther et al. (2010) ⁵²	UK	Single	15	68 med	n/a	Ischaemic CMP	30	CARTO 3v4 (Biosense Webster, Diamond Bar, CA, USA)	Ripple, voltage, point by point or using an automated point collection facility (CONFIDENSE Continuous mapping)	85% (non-inducible in 2/15 at the beginning)	None	Multielectrode Pentaray catheter; 3.5 mm tip SmartTouch ThermoCool catheter (Biosense Webster)	6 med	71	Endo	n/a	
Morai et al. (2010) ⁵³	Israel	Single	11	71	100	Ischaemic CMP	23	CARTO mapping and navigation system (Biosense Webster, Johnson & Johnson, USA)	Pace and activation, voltage and substrate	82% (termination of clinical VT and/or non-inducibility)	n/a	Open-irrigated ablation catheter	3	11.1	Endo	n/a	
Volkmer et al. (2006) ⁵⁴	EU	Single	47	65	91.5	Ischaemic CMP	30	CARTO	Activation, voltage, substrate, entrainment and pace	79.1% (non-inducibility of the clinical or any slower VT)	n/a	4 mm tip (Navistar™) or a 3.5 mm irrigated tip electrode (Navistar ThermoCool™, Biosense Webster Ltd), ReStar™, (Biosense Webster Ltd)	25.1	43.2	Endo	23.7±19.2	
Anz et al. (2007) ⁵⁵	EU	Single	69	66.2	88.4	Ischaemic CMP	32.6	CARTO system (Biosense Webster, Inc)	Voltage, entrainment and pace	90% (non-inducibility of the clinical or slower VT), 43 (63%) (non-inducibility of any VT)	n/a	Quadrupolar 6F Josephson catheter (Biosense Webster, Inc), decapolar 6F Parahis catheter (Biosense Webster, Inc), 7F-CARTO catheter, either non-irrigated or irrigated (3.5 mm tip; Navistar, Biosense Webster Ltd.) or irrigated (3.5 mm tip; Navistar ThermoCool, Biosense Webster Ltd.)	25	37.7	Endo	24±13.5	
Bogun et al. (2005) ⁵⁶	US	Single	23	68	n/a	Ischaemic CMP	21	CARTO Biosense Webster Inc.	Voltage map	74% of VTs (targeted VT was successfully ablated)	n/a	Navistar catheter (Biosense Webster Inc.)	n/a	n/a	Endo	n/a	
Brunckhorst et al. (2004) ⁵⁷	US	Single	11	68	100	Ischaemic CMP	24	CARTO (Biosense Cordis Webster)	Voltage, pace and entrainment	90.9% successful ablation sites were localized within an isthmus identified by pace mapping) 45.5% (absence of inducible VT) 45.5% (absence of inducible clinical VT)	n/a	n/a	n/a	n/a	Endo	n/a	
Denkge et al. (2005) ⁵⁸	EU	Single	25	62	n/a	Ischaemic CMP	37	CARTOTM system (Biosense Webster®)	Substrate map, voltage mapping and pace	70% (no VT inducible) procedural success (epicardial origin-problematic access to the LV) (n=2)	Patients without procedural success (epicardial origin-problematic access to the LV) (n=2)	Navistar DS catheter	10	17.4	Endo	n/a	

Continued

Table 2: Continued

Study (year)	Country	Single/ multicentre	N	Mean age (years)	Males (%)	Clinical setting	EF (%)	Mapping system	Mapping type	Acute success/ definition of success	Complications (n)	Catheter	Follow up (months)	Recurrence at follow up (%)	Endo or epi or both	Fluoroscopy time (min) mean±SD
Dinov et al. (2012) ³⁹	EU	Single	102	67.7	86.3	Ischaemic CMP	32	Niobe Stereotaxis magnetic navigation system (Stereotaxis Inc.), electroanatomical mapping system (CARTO-RMT; Biosense Webster, Inc.) or EnSite- NavX (St. Jude Medical Inc., St. Paul, MN, USA)	Substrate and/ or activation mapping entrainment pace	96% (successful ablation of the clinical VT) 76.5% (successful ablation of all inducible monomorphic sustained VTs)	Pseudoaneurysm (n=2), arteriovenous fistula (n=1), pericardial effusions (n=6) and liver injury (n=1)	Irrigated-tip catheter (Navistar ThermoCool; Biosense Webster Inc.) and irrigated-tip catheter (Navistar ThermoCool-RMT; Biosense Webster Inc. or Tigrum Flux Gold, Biotronik, Berlin, Germany)	14 med	26.9	Endo-epi in one patient	n/a
Kettinger et al. (2010) ³⁸	EU	Single	7	66.6	n/a	Ischaemic CMP	32.1	CARTO system (Biosense Webster)	Voltage and pace	100% (elimination of all clinically documented or inducible VTs)	No major complications	6F quadripolar diagnostic catheter (Biosense Webster), irrigated-tip ablation catheter (NAVI-STAR, 7F; D- (or C-) type, 3.5 mm- tip; Biosense Webster) and 8 mm-tip ablation catheter (7F; mostly C-curve; Biosense Webster)	19	28.6	Endo	n/a
Li et al. (2009) ⁴¹	EU and China	Single	14	65	92.9	Ischaemic CMP	29	CARTO, NAVI-STAR (Cordis-Webster, Johnson and Johnson)	Substrate, voltage, pace and entrainment	78.6% (no VT inducibility)	n/a	Navistar, ThermoCool (Cordis-Webster)	7	21.4	Endo	10±7
Jamil-Copley et al. (2015) ⁴²	EU	Single	21	69	95	Ischaemic CMP	28	CARTO-3 (Biosense Webster Inc.)	Voltage, LAT, entrainment, ripple	19.1% (termination during ablation) 78% (non-inducibility)	n/a	5 mm Navistar ThermoCool catheter (Biosense Webster Inc.)	15.5	29	Endo	n/a
Verma et al (2005) ⁴³	USA	Single	46	65	89.1	Ischaemic CMP	27	CARTO (Biosense Webster, Inc.)	Activation, voltage and entrainment	100% (identification of the successful ablation sites)	Periprocedural stroke (n=1), femoral pseudoaneurysm (n=1) and prolonged hypotension (n=1)	7F Navistar (Biosense Webster, Inc.)	17 med	36.9	Endo-epi	n/a
Miyamoto et al. (2015) ⁴⁴	Japan	Single	11	59	24	Ischaemic, dilated CMP, sarcoidosis and HCM	24	CARTO version 3 or XP	Voltage, activation, entrainment and pace	45.5% (VT termination and non-inducibility)	n/a	Open-irrigated tipped catheter, ThermoCool® (Biosense Webster, Johnson & Johnson)	21	-	Endo-epi	n/a
Yamashina et al. (2009) ⁴⁴	Japan	Single	72	43.6	44.4	Idiopathic RVOT	n/a	CARTO system (Biosense Webster)	Activation, voltage and pace	63 (87.5%) (absence of any spontaneous or induced clinical RVOT arrhythmias and no recurrence of any symptomatic ventricular arrhythmia)	n/a	7 F Navistar (Biosense Webster)	n/a	n/a	Endo	n/a
Suleiman et al. (2007) ¹⁵	Israel	Single	15	59	49	RVOT VT	n/a	CARTO	-	12 (80)	None		33* refers to the total population	26.7	-	10.3* refers to the total population
Parrera et al. (2013) ⁴⁵	EU	Single	32	43	25	Outflow tract VT	n/a	CARTO XP RMT (Biosense Webster)	Pace mapping	26 (81%) (first procedure) (suppression and non-inducibility of arrhythmia)	None	Navistar RMT (Biosense Webster), Navistar RMT ThermoCool (Biosense Webster)	10.2	6.3	Endo	10±7.8

Continued

Table 2: Continued

Study (year)	Country	Single/ multicentre	N	Mean age (years)	Males (%)	Clinical setting	EF (%)	Mapping system	Mapping type	Acute success/ definition of success	Complications (n)	Catheter	Follow up (months)	Recurrence at follow up (%)	Endo or epi or both	Fluoroscopy time (min) mean±SD
Yamashina et al. (2016) ^{62*}	Japan	Single	33	45.5	36.4	Outflow tract VT	n/a	CARTO (Biosense Webster)	Activation and pace	87.9% (absence of any spontaneous or induced clinical RVOT arrhythmias and no recurrence of any symptomatic ventricular arrhythmia)	None	Navistar (Biosense Webster) catheter	2	6.1%	Endo	n/a
Tovia-Brodie et al. (2016) ⁶⁸	Israel	Single	18	60	66.7	Idiopathic (77.8%) and ischaemic VT (22.2%) (complete abolition of the clinical ventricular arrhythmia or non-inducibility of ischaemic ventricular tachycardia)	48.6	CARTO Segmentation Module software (Biosense Webster)	Activation, pace and substrate mapping	Idiopathic: 78.6% and ischaemic: 100%	None	3.5 mm open-irrigated catheter (ThermoCool, SmartTouch; Biosense Webster)	n/a	n/a	Endo and epi for ischaemic VT	Ischaemic: 20.9±8.8 min and idiopathic: 22.95±12 min
Verma et al. (2005) ⁶⁹	USA	Single	22	41	68	ARVC	55	CARTO mapping system (Biosense Webster Inc.)	Voltage and pace	18 (82%) (non-inducibility of ablated VT or other sustained monomorphic VT)	Cardiac tamponade (1) and femoral haematoma (2)	7F 4 mm-tip deflectable ablation catheter (Navistar; Biosense Webster Inc.)	37 med	36.4	Endo	83±47
Satomi et al. (2006) ⁷⁰	Japan	Single	17	47	76.5	ARVC	n/a	CARTO	Activation, voltage and pace	88% [†] (no monomorphic VT was inducible)	No major complications	7F mapping/ablation catheter (NAVI-STAR; Cordis-Webstar, Johnson & Johnson)	26	23.5	Endo	n/a
Rhythmia mapping system																
Martin et al. (2019) ⁷¹	EU	Multicentre	27	64.3	85.2	Ischaemic and dilated CMP	36	Rhythmia	Substrate, activation, entrainment and pace	92.3% (non-inducibility)	No complications	Orion mapping catheter (Boston Scientific)	51.6	18.5	Endo-epi	44.4±12.4
Nijmich et al. (2017) ⁷²	EU	Single	22	67	n/a	Ischaemic and non-ischaemic, ARVC	36	Rhythmia 3D electro-anatomical mapping system (Boston Scientific, Marlborough, MA, USA)	Voltage and activation	87% (non-inducibility)	Tamponade (n=1)	ThermoCool, D- or F-Type, 2–5–2 mm spacing (Biosense Webster) or Intella NAV Or (Boston Scientific)	4	10	Endo-epi	20.7±1.6
Sultan et al. (2019) ⁷³	EU	Single	32 (VT 15 patients and VE 17 patients)	63	90.6	Different clinical settings including ischaemic CMP and myocarditis	47.2	IntellaMap Orion™ Mapping Catheter (Boston Scientific Corporation) in combination with the Rhythmia™ mapping system (Boston Scientific Corporation)	Activation, pace, substrate and entrainment	100% (no inducibility)	Femoral haematomas (n=3)	Quadripolar diagnostic catheter (Inquiry™, 5 F; Fa.; Abbott), decapolar diagnostic catheter (Inquiry™, 6 F; Fa.; Abbott)	6	20	Endo	23.4±13.7
Viswanathan et al. (2016) ⁷⁴	EU	Single	19 (VT: 12 patients and VE: 7 patients)	64	79	Ischaemic and non-ischaemic CMP, congenital HD, hypertrophic CMP and normal	35	Rhythmia (Boston Scientific Inc.)	Activation	VE: 67 (86%) VT procedures: 8/13 (61.5%) complete success (non-inducibility)	Femoral pseudoaneurysm (n=1) and groin haematoma (n=1)	64-electrode mini-basket mapping catheter (IntellaMap Orion™; Boston Scientific Inc.)	10	VT 25% VE: 14.3%	Endo-epi	41.8±17.5
NavX-EnSite-Cardioinsight mapping systems																
Lee et al. (2019) ⁷⁵	Taiwan	Single	28 (24 PVC and 4 VTs)	48.8	42.9	RVOT PVCs or VT	70	EnSite NavX or Velocity V5.0 3D mapping system (Abbott)	Voltage, pace and activation	78.6% (at least an 80% decrease in PVC burden by 24 h Holter ECGs 3 months later after ablation)	1% pericardial effusion	7F 4 mm quadripolar irrigated ablation catheter or a 7F 4 mm or 8 mm non-irrigated ablation catheter	3	21.4	Endo	n/a

Continued

Table 2: Continued

Study (year)	Country	Single/ multicentre	N	Mean age (years)	Males (%)	Clinical setting	EF (%)	Mapping system	Mapping type	Acute success/ definition of success	Complications (n)	Catheter	Follow up (months)	Recurrence at follow up (%)	Endo or epi or both	Fluoroscopy time (min) mean±SD
Nayyar et al. (2013) ⁶	Australia	Single	22	67	95	Ischaemic CMP	32	EnSite NavX (St. Jude Inc.)	Entrainment and pace	64% (no inducible VT after ablation, abolition of ≥1 clinical VTs with other VTs remaining inducible was considered a partial success and the inability to eliminate the clinical VT was considered as a failure)	n/a	3.5 mm tip irrigated ablation (Coolflex (St. Jude) or ThermoCool (Biosense Webster), 20-pole catheter (Penafax), 2–6 mm inter-electrode spacing and 1 mm electrodes (Biosense Webster, Inc.)	16	22.7	Endo	n/a
Miyamoto et al. (2010) ⁷	Japan	Single	55	52	49%	Organic heart disease: 7 patients and idiopathic VT: 48 patients	n/a	EnSite version 3.0 in 20 patients and version 6.0i in 35 patients	Voltage, substrate, activation and entrainment	95% (sustained VT was VT termination, subsequent non- inducibility of VT for focal VT and non-inducibility for non-sustained focal VT and PVC)	No complications	Quadrupolar electrode catheter, 20-pole multielectrode catheters (St. Jude Medical, Minnetonka, MN, USA/Ten-Ten, St. Jude Medical)	21	-	Endo	30±21
Nair et al. (2011) ⁸	India	Single	15	44	80	ARVC	n/a	EnSite array mapping and non-contact electroanatomical mapping	Activation, entrainment and pace	86.7% (all of the inducible VTs were successfully mapped and ablated)	No complications	-	25	13.3	Endo	n/a
Hocini et al. (2015) ⁹	EU	Multicentre	24	45	58	Idiopathic PVCs and PVCs in the setting of HCM and ischaemic CMP	58.4%	3D mapping technique (ECVUE; Cartoonsight Inc.)	Activation	100%	None	Quadrupolar mapping catheter) and 4 mm tip ablation catheter (Biosense Webster)	24.7	4.2	Endo	8.8±1.5

*These were similar cohorts and, therefore, were included once in the quantitative synthesis.

⁶The percentage refers to the number of successfully ablated VTs.

ARVC = arrhythmogenic right ventricular cardiomyopathy; CMP = cardiomyopathy; EF = ejection fraction; endo = endocardial; Epi = epicardial; F = French; HCM = hypertrophic cardiomyopathy; med = median; n/a = not available; PVCs = premature ventricular complexes; RVOT = right ventricular outflow tachycardia; VE = ventricular ectopy; VT = ventricular tachycardia.

Figure 1: Flow diagram of the search strategy.





Figure 2: Acute success rates and complications of commonly used mapping systems



AF = atrial fibrillation; AT = atrial tachycardia; CMP = cardiomyopathy; VT = ventricular tachycardia

studies) CA ablation procedures.^{52–63} Specifically, the acute success rates are defined as the absence of VT inducibility at the end of the procedure, which ranged from 70 to 100%.^{52–63}

Furthermore, the CARTO mapping system showed excellent results with more than 80% acute success rate in outflow tract ventricular arrhythmias and in the setting of ARVC (four and two studies, respectively).^{15,65–67,69,70}

Rhythmia mapping system

The search strategy revealed four studies on the acute arrhythmia termination rates using the Rhythmia mapping system in the setting of ventricular arrhythmias (*Table 2*). These studies showed excellent results in abolishing VT and ventricular ectopy in different clinical settings, including ischaemic and non-ischaemic cardiomyopathy, with acute success rates ranging from 61.5 to 100%.^{71–74}

EnSite/NavX mapping systems

The search strategy revealed five studies on the success rates of the EnSite/NavX mapping system, mainly in the setting of patients with outflow tract ventricular arrhythmias, ischaemic cardiomyopathy and ARVC (*Table 2*).^{75–79} Specifically, acute success rates were found to be 78.6% in patients with outflow tract arrhythmias, 86.7% in patients with ARVC and 64% in patients with ischaemic cardiomyopathy.^{75–79}

Discussion

Our review summarizes the acute success rates of the most popular mapping systems across different clinical settings. The existing evidence shows that the most commonly used mapping systems have excellent efficacy regarding acute arrhythmia termination outcomes in both atrial and ventricular arrhythmias, depending mainly on the type of arrhythmia and the clinical setting.

Cardiac mapping is an essential component in the understanding and treatment of arrhythmias through CA procedures. Activation and electrocardiographic signal amplitude mapping are the most frequent modalities in 3D mapping systems, while entrainment mapping can provide additional data in the setting of atrial or ventricular arrhythmias. The commonly used 3D mapping systems such as CARTO (Biosense Webster), EnSite Precision (Abbott) and, more recently, Rhythmia (Boston Scientific) systems have played a major role in enabling and facilitating the high-density mapping of complex arrhythmias.³ The *in vitro* and *in vivo* accuracy of the CARTO mapping system has been studied since 1997.⁸⁴ The CARTO mapping system has been used to guide the ablation procedure in patients with different types of atrial arrhythmias, including AF, AFL and AT.^{11–14,16–26,82}

Today, the CARTO 3 system consists of a location pad with three separate low-level magnetic field-emitting coils arranged as a triangle under the patient's body and six-electrode patches positioned on the patient's back and chest. The latest version relies on hybrid magnetic- and current-based localization technologies and has reported an accuracy of less than 1 mm.³ On the other hand, the current EnSite/NavX navigation and visualization technology consists of a set of three pairs of skin patches and a system reference patch, and uses hybrid impedance-based and magnetic-tracking technologies to create cardiac models or geometries, which display voltage data and activation timing. The Rhythmia mapping system, which uses a hybrid tracking technology using both magnetic- and impedance-based localization features for map creation, was developed as a high-definition system incorporating a high-resolution 64-electrode catheter, 3D ultra-high-density mapping capabilities and an algorithm that automates the signal capture and mapping processes.^{3,85}

Each of the mentioned systems, except NavX, uses its catheters for high-density mapping. CARTO 3 uses the PentaRay catheter and the Rhythmia system uses the IntellaMap Orion catheter, while the NavX system allows the operator to use any available catheter.

Single and multielectrode acquisition techniques are clinically effective, whereas a trend towards multielectrode use has been observed in the mapping of complex arrhythmias, such as left AFL and scar-related VT.⁸⁶ A pooled analysis of studies that compared fluoroscopy- with non-fluoroscopy-guided ablation of AF and AFL showed that non-fluoroscopic mapping and navigational systems yielded acute success rates not significantly different from fluoroscopy-guided ablation, while navigational systems resulted in a 10% relative reduction in the overall failure rate compared with fluoroscopy-guided ablation for the treatment of AF.⁸⁷ Another interesting finding was that mapping systems were shown to reduce the arrhythmia burden and the need for antiarrhythmic drugs in patients with complex arrhythmia following a failed fluoroscopy-guided ablation procedure.⁸⁷

Few studies in the literature have performed direct comparisons of the most commonly used mapping systems. Specifically, Liu et al. performed a direct comparison of CARTO-guided with EnSite/NavX-guided ablation of the PV in AF.⁸⁸ They found that compared with the CARTO group, the EnSite/NavX group had a significantly higher acute success rate. Specifically, AF was terminated by radiofrequency delivery in 14 cases (35%) using EnSite/NavX system versus 5 cases (14%) in the CARTO system ($p < 0.05$).⁸⁸ In addition, complete PV isolation was achieved in 26 cases (65%) in the EnSite/NavX group versus 11 cases (31%) in the CARTO group ($p < 0.05$), although contemporary techniques and outcomes have resulted in much higher success rates.⁸⁸ On the other hand, Choo et al. did not find a significant difference in acute success rates between CARTO and NavX mapping systems in the setting of paroxysmal or persistent AF CA.¹² In another study, the acute outcomes of AT ablation using standard (CARTO™ and NavX™) versus Rhythmia™ 3D high-density mapping systems were compared; in this setting, acute success rates were found to be similar for any system, leading to around 75% complete and 93% partial acute success rates in a highly selected population.⁸⁹ Rottner et al. performed a direct comparison of CARTO and Rhythmia mapping systems in the setting of AF.⁹⁰ This study showed that the Rhythmia mapping system had a significantly longer total mapping time, a longer total fluoroscopy time, more delivered RF applications and a longer total RF duration compared with the CARTO system, while there was no difference regarding the total ablation time, total procedure duration and acute procedural success.⁹⁰ Another study evaluated the effect of Rhythmia in terms of the outcome of the second ablation for AF compared with the conventional method with an additional anatomical guide by the CARTO system.⁹¹ The authors have found that high-density mapping for the second ablation of AF was superior to the conventional ablation method in terms of the suppression of atrial events.⁹¹ Kaseno et al. evaluated PentaRay®/CARTO® 3 and Orion™/Rhythmia™ in LA voltage mapping.⁹² The study showed that the PentaRay map had a shorter procedure time than the Orion map, while a discrepancy in the evaluation of low-voltage areas between PentaRay and Orion maps was revealed.⁹²

Finally, it should also be noted that while acute AF termination has been a common goal of AF ablation, in more recent years, the focus is more on a substrate ablation strategy (PVI and often additional ablation), and now the acute completeness of PVI and long-term freedom from arrhythmias are more typical endpoints.⁵

Limitations

We identified only a small number of studies^{27–51} on acute arrhythmia termination during CA procedures for the Rhythmia and NavX/EnSite mapping systems, especially in the setting of ventricular arrhythmias. Specifically, no data were revealed by the search strategy about the acute success rates in AF CA procedures using the Rhythmia mapping system, while only one study (Nayyar et al.) provided data about ischaemic VT for the EnSite mapping system.⁷⁶ Furthermore, no data on the CardiInsight mapping system were retrieved.

It should be noted that Rhythmia is a more recently developed mapping system compared with CARTO and EnSite; therefore, the studies on Rhythmia included in this analysis are likely to be more recent, compared with studies on CARTO and EnSite and, therefore, could skew the results.

A quantitative synthesis to estimate the pooled success rate of each mapping system was not performed. With regard to AF, beyond PVI, there is no single established strategy; as a result, differences in success rates are also dependent on the ablation strategy. Furthermore, the type of AF (paroxysmal, persistent and long-standing persistent AF) also influences the success rate of the ablation procedure. Similarly, AT ablation is largely dependent on the underlying arrhythmia mechanism, atrial substrate and ablation strategy. With regard to the ventricular arrhythmias, both the type of cardiomyopathy and ablation strategy can influence the ablation outcomes. Ablation for VT is often performed during SR using

substrate modification of local abnormal ventricular activities due to the hemodynamic instability associated with the arrhythmia.⁹³ Therefore, our results cannot be extrapolated to procedures using substrate modification for ventricular arrhythmias.

We chose not to provide data about the procedural time presented in each study because this parameter is dependent on not only the technical characteristics of each system but also the operator's skills. In addition, this review did not focus on the long-term arrhythmia recurrence rate because this marker is highly dependent on patients' comorbidities, post-procedural medications, echocardiographic findings, follow-up strategy for the identification of arrhythmia recurrence, etc. It should be noted that acute success definitions differ slightly among the included studies (*Tables 1 and 2*), and this is an additional limitation that prevents the comparison of the efficacy of the different mapping systems in all clinical settings. Another cause of the noted discrepancy in success rates for each mapping system among the included studies could be related to the new features that were incorporated into each system and the different catheters that were used in the last few years.

Conclusions

Mapping systems have played a crucial role in high-density mapping and the observed high procedural success rates of atrial and ventricular CA procedures. More data are needed about the comparative efficacy of the different mapping systems across different clinical settings. □

- Hindricks G, Potpara T, Dagres N, et al. ESC Guidelines for the diagnosis and management of atrial fibrillation developed in collaboration with the European Association for Cardio-Thoracic Surgery (EACTS): The task force for the diagnosis and management of atrial fibrillation of the European Society of Cardiology (ESC) developed with the special contribution of the European Heart Rhythm Association (EHRA) of the ESC. *Eur Heart J*. 2021;37:493–498. DOI: 10.1093/eurheartj/ehaa798.
- Cronin EM, Bogun FM, Maury P, et al. HRS/EHRA/APHRS/LAHS expert consensus statement on catheter ablation of ventricular arrhythmias. *Heart Rhythm*. 2020;17:e2–154. DOI: 10.1016/j.hrthm.2019.03.002.
- Kim Y-H, Chen S-A, Ernst S, et al. APHRS expert consensus statement on three-dimensional mapping systems for tachycardia developed in collaboration with HRS, EHRA, and LAHRS. *J Arrhythm*. 2020;36:215–70. DOI: 10.1002/joa3.12308.
- Calikins H, Brugada J, Packer DL, et al. HRS/EHRA/ECAS expert consensus statement on catheter and surgical ablation of atrial fibrillation: recommendations for personnel, policy, procedures and follow-up. A report of the Heart Rhythm Society (HRS) Task Force on Catheter and Surgical Ablation of Atrial Fibrillation developed in partnership with the European Heart Rhythm Association (EHRA) and the European Cardiac Arrhythmia Society (ECAS); in collaboration with the American College of Cardiology (ACC), American Heart Association (AHA), and the Society of Thoracic Surgeons (STS). Endorsed and approved by the governing bodies of the American College of Cardiology, the American Heart Association, the European Cardiac Arrhythmia Society, the European Heart Rhythm Association, the Society of Thoracic Surgeons, and the Heart Rhythm Society. *Europace*. 2007;9:335–79. DOI: 10.1093/europace/eum120.
- Katritsis D, Merchant FM, Mela T, et al. Catheter ablation of atrial fibrillation: the search for substrate-driven end points. *J Am Coll Cardiol*. 2010;55:2293–8. DOI: 10.1016/j.jacc.2010.03.016.
- Hunter RJ, Berriman TJ, Diab I, et al. Long-term efficacy of catheter ablation for atrial fibrillation: Impact of additional targeting of fractionated electrograms. *Heart*. 2010;96:1372–8. DOI: 10.1136/hrt.2009.188128.
- Calikins H, el-Atassi R, Kalbfleisch SJ, et al. Effect of operator experience on outcome of radiofrequency catheter ablation of accessory pathways. *Am J Cardiol*. 1993;71:1104–5. DOI: 10.1016/0002-9149(93)90581-v.
- Bazoukis G, Letsas KP, Tse G, et al. Predictors of arrhythmia recurrence in patients with heart failure undergoing left atrial ablation for atrial fibrillation. *Clin Cardiol*. 2018;41:63–7. DOI: 10.1002/clc.22850.
- Efremidis M, Letsas KP, Georgopoulos S, et al. Safety, long-term outcomes and predictors of recurrence following a single catheter ablation procedure for atrial fibrillation. *Acta Cardiol*. 2019;74:319–24. DOI: 10.1080/00015385.2018.1494114.
- Moher D, Liberati A, Tetzlaff J, et al. Preferred reporting items for systematic reviews and meta-analyses: The PRISMA statement. *PLoS Med*. 2009;6:e1000097. DOI: 10.1371/journal.pmed.1000097.
- Calvo D, Rubin J, Pérez D, Moris C. Ablation of rotor domains effectively modulates dynamics of human: Long-standing persistent atrial fibrillation. *Circ Arrhythm Electrophysiol*. 2017;10:12. DOI: 10.1161/CIRCEP.117.005740.
- Choo WK, Farwell D, Harris S. Experience of atrial fibrillation ablation in a new cardiac centre using three-dimensional mapping and multielectrode duty-cycled radiofrequency ablation. *Arch Cardiovasc Dis*. 2011;104:396–402. DOI: 10.1016/j.acvd.2011.05.003.
- Nademanee K, McKenzie J, Kosar E, et al. A new approach for catheter ablation of atrial fibrillation: Mapping of the electrophysiologic substrate. *J Am Coll Cardiol*. 2004;43:2044–53. DOI: 10.1016/j.jacc.2003.12.054.
- Seitz J, Horvilleur J, Lacotte J, et al. Automated detection of complex fractionated atrial electrograms in substrate-based atrial fibrillation ablation: Better discrimination with a new setting of CARTO® algorithm. *J Atr Fibrillation*. 2013;6:673. DOI: 10.4022/jafib.673.
- Suleiman M, Gepstein L, Roguin A, et al. Catheter ablation of complex fractionated atrial electrograms by electroanatomic imaging (CARTO): A single-center experience. *Isr Med Assoc J*. 2007;9:260–4.
- Takahashi Y, Yamashita S, Suzuki M, et al. Efficacy of catheter ablation of focal sources in persistent atrial fibrillation. *J Cardiovasc Electrophysiol*. 2018;29:559–65. DOI: 10.1111/jce.13415.
- Wang X, Liu X, Shi H, et al. Heart rhythm disorders and pacemakers: Pulmonary vein isolation combined with substrate modification for persistent atrial fibrillation treatment in patients with valvular heart diseases. *Heart*. 2009;95:1773–83. DOI: 10.1136/hrt.2007.124594.
- Wnuk-Wojnar A-M, Trusz-Gluzka M, Czerwiński C, et al. Circumferential pulmonary vein RF ablation in the treatment of atrial fibrillation: 3-year experience of one centre. *Kardiol Pol*. 2005;63:362–70.
- Wu J, Estner H, Luik A, et al. Automatic 3D mapping of complex fractionated atrial electrograms (CFAE) in patients with paroxysmal and persistent atrial fibrillation. *J Cardiovasc Electrophysiol*. 2008;19:897–903. DOI: 10.1111/j.1540-8167.2008.01145.x.
- Esato M, Hindricks G, Sommer P, et al. Color-coded three-dimensional entrainment mapping for analysis and treatment of atrial macroreentrant tachycardia. *Heart Rhythm*. 2009;6:349–58. DOI: 10.1016/j.hrthm.2008.12.013.
- Jamil-Copley S, Linton N, Koa-Wing M, et al. Application of ripple mapping with an electroanatomic mapping system for diagnosis of atrial tachycardias. *J Cardiovasc Electrophysiol*. 2013;24:1361–9. DOI: 10.1111/jce.12259.
- Okumura K, Sasaki S, Kimura M, et al. Usefulness of combined CARTO electroanatomical mapping and manifest entrainment in ablating adenosine triphosphate-sensitive atrial tachycardia originating from the atrioventricular node vicinity. *J Arrhythm*. 2016;32:133–40. DOI: 10.1016/j.joa.2015.11.004.
- Strisciuglio T, Vandersickel N, Lorenzo G, et al. Prospective evaluation of entrainment mapping as an adjunct to new-generation high-density activation mapping systems of left atrial tachycardias. *Heart Rhythm*. 2020;17:211–9. DOI: 10.1016/j.hrthm.2019.09.014.
- Vicera JJB, Lin Y-J, Lee P-T, et al. Identification of critical isthmus using coherent mapping in patients with scar-related atrial tachycardia. *J Cardiovasc Electrophysiol*. 2020;31:1436–47. DOI: 10.1111/jce.14457.
- Wo H-T, Wen M-S, Chang P-C, et al. Successful treatment of macroreentrant atrial tachycardia by radiofrequency ablation targeting channels with continuous activation. *Pacing Clin Electrophysiol*. 2014;37:927–37. DOI: 10.1111/pace.12408.
- Yagishita A, Takahashi Y, Kawabata M, et al. Utility of a ripple map for the interpretation of atrial propagation during atrial tachycardia. *J Interv Card Electrophysiol*. 2019;56:249–57. DOI: 10.1007/s10840-019-00638-4.
- Kitamura T, Martin R, Denis A, et al. Characteristics of single-loop macroreentrant biatrial tachycardia diagnosed by ultrahigh-resolution mapping system. *Circ Arrhythm Electrophysiol*. 2018;11:e005558. DOI: 10.1161/CIRCEP.117.005558.
- Yamashita S, Takigawa M, Denis A, et al. Pulmonary vein-gap re-entrant atrial tachycardia following atrial fibrillation ablation: An electrophysiological insight with high-resolution mapping. *Europace*. 2019;21:1039–47. DOI: 10.1093/europace/euz034.
- Anter E, McElderry TH, Contreras-Valdes FM, et al. Evaluation of a novel high-resolution mapping technology for ablation of recurrent scar-related atrial tachycardias. *Heart Rhythm*. 2016;13:2048–55. DOI: 10.1016/j.hrthm.2016.05.029.
- De Simone A, Anselmino M, Scaglione M, et al. Is the mid-diastolic isthmus always the best ablation target for re-entrant atrial tachycardias? *J Cardiovasc Med (Hagerstown)*. 2020;21:113–22. DOI: 10.2459/JCM.0000000000000923.
- Lařcu DG, Bun S-S, Viera F, et al. Selection of critical isthmus in scar-related atrial tachycardia using a new automated ultrahigh resolution mapping system. *Circ Arrhythm Electrophysiol*. 2017;10:e004510. DOI: 10.1161/CIRCEP.116.004510.
- Takigawa M, Derval N, Frontera A, et al. Revisiting anatomic macroreentrant tachycardia after atrial fibrillation ablation using ultrahigh-resolution mapping: Implications for ablation. *Heart Rhythm*. 2018;15:326–33. DOI: 10.1016/j.hrthm.2017.10.029.
- Takigawa M, Derval N, Maury P, et al. Comprehensive multicenter study of the common isthmus in post-atrial fibrillation ablation multiple-loop atrial tachycardia. *Circ Arrhythm Electrophysiol*. 2018;11:e006019. DOI: 10.1161/CIRCEP.117.006019.
- Aksu T, Guler TE, Bozyel S, et al. Initial experience with fractionation mapping-guided ablation strategy in patients with long-standing persistent atrial fibrillation. *J Interv Card Electrophysiol*. 2021;61:405–13. DOI: 10.1007/s10840-020-00834-7.
- Miyamoto K, Tsuchiya T, Narita S, et al. Bipolar Electrogram amplitudes in the left atrium are related to local conduction velocity in patients with atrial fibrillation. *Europace*. 2009;11:597–605. DOI: 10.1093/europace/eup352.
- Miyamoto K, Tsuchiya T, Nagamoto Y, et al. Characterization of bipolar electrograms during sinus rhythm for complex

- fractionated atrial electrograms recorded in patients with paroxysmal and persistent atrial fibrillation. *Europace*. 2010;12:494–501. DOI: 10.1093/europace/euq033.
37. Yamaguchi T, Tsuchiya T, Miyamoto K, et al. Characterization of non-pulmonary vein foci with an EnSite array in patients with paroxysmal atrial fibrillation. *Europace*. 2010;12:1698–706. DOI: 10.1093/europace/euq326.
 38. Ammar-Busch S, Reents T, Knecht S, et al. Correlation between atrial fibrillation driver locations and complex fractionated atrial electrograms in patients with persistent atrial fibrillation. *Pacing Clin Electrophysiol*. 2018;41:1279–85. DOI: 10.1111/pace.13483.
 39. Kumagai K, Sakamoto T, Nakamura K, et al. Combined dominant frequency and complex fractionated atrial electrogram ablation after circumferential pulmonary vein isolation of atrial fibrillation. *J Cardiovasc Electrophysiol*. 2013;24:975–83. DOI: 10.1111/jce.12166.
 40. Kumagai K, Minami K, Kutsuzawa D, Oshima S. Evaluation of the characteristics of rotational activation at high-dominant frequency and complex fractionated atrial electrogram sites during atrial fibrillation. *J Arrhythm*. 2017;33:49–55. DOI: 10.1016/j.joa.2016.05.008.
 41. Lin YJ, Tai CT, Chang SL, et al. Efficacy of additional ablation of complex fractionated atrial electrograms for catheter ablation of nonparoxysmal atrial fibrillation. *J Cardiovasc Electrophysiol*. 2009;20:607–15. DOI: 10.1111/j.1540-8167.2008.01393.x.
 42. Lo L-W, Lin Y-J, Tsao H-M, et al. The impact of left atrial size on long-term outcome of catheter ablation of chronic atrial fibrillation. *J Cardiovasc Electrophysiol*. 2009;20:1211–6. DOI: 10.1111/j.1540-8167.2009.01546.x.
 43. Matsuo S, Yamane T, Date T, et al. Substrate modification by pulmonary vein isolation and left atrial linear ablation in patients with persistent atrial fibrillation: Its impact on complex fractionated atrial electrograms. *J Cardiovasc Electrophysiol*. 2012;23:962–70. DOI: 10.1111/j.1540-8167.2012.02322.x.
 44. Nair M, Nayyar S, Rajagopal S, et al. Results of radiofrequency ablation of permanent atrial fibrillation of >2 years duration and left atrial size >5 cm using 2-mm irrigated tip ablation catheter and targeting areas of complex fractionated atrial electrograms. *Am J Cardiol*. 2009;104:683–8. DOI: 10.1016/j.amjcard.2009.04.042.
 45. Nakahara S, Hori Y, Kobayashi S, et al. Epicardial adipose tissue-based defragmentation approach to persistent atrial fibrillation: Its impact on complex fractionated electrograms and ablation outcome. *Heart Rhythm*. 2014;11:1343–51. DOI: 10.1016/j.hrthm.2014.04.040.
 46. Roux J-F, Gojraty S, Bala R, et al. Effect of pulmonary vein isolation on the distribution of complex fractionated electrograms in humans. *Heart Rhythm*. 2009;6:156–60. DOI: 10.1016/j.hrthm.2008.10.046.
 47. Suenari K, Lin Y-J, Chang S-L, et al. Relationship between arrhythmogenic pulmonary veins and the surrounding atrial substrate in patients with paroxysmal atrial fibrillation. *J Cardiovasc Electrophysiol*. 2011;22:405–10. DOI: 10.1111/j.1540-8167.2010.01932.x.
 48. Verma A, Lakkireddy D, Wulffhart Z, et al. Relationship between complex fractionated electrograms (CFE) and dominant frequency (DF) sites and prospective assessment of adding DF-guided ablation to pulmonary vein isolation in persistent atrial fibrillation (AF). *J Cardiovasc Electrophysiol*. 2011;22:1309–16. DOI: 10.1111/j.1540-8167.2011.02128.x.
 49. Narita S, Miyamoto K, Tsuchiya T, et al. Radiofrequency catheter ablation of atrial tachycardia under navigation using the EnSite array. *Circ J*. 2010;74:1322–31. DOI: 10.1253/circj.cj-09-1008.
 50. Patel AM, d'Avila A, Neuzil P, et al. Atrial tachycardia after ablation of persistent atrial fibrillation: Identification of the critical isthmus with a combination of multielectrode activation mapping and targeted entrainment mapping. *Circ Arrhythm Electrophysiol*. 2008;1:14–22. DOI: 10.1161/CIRCEP.107.748160.
 51. Nagamoto Y, Tsuchiya T, Miyamoto K, et al. Atrial tachycardia during ongoing atrial fibrillation ablation. - EnSite array analysis. *Circ J*. 2011;75:1080–9. DOI: 10.1253/circj.cj-10-0742.
 52. Luther V, Linton NWF, Jamil-Copley S, et al. A prospective study of ripple mapping the post-infarct ventricular scar to guide substrate ablation for ventricular tachycardia. *Circ Arrhythm Electrophysiol*. 2016;9:e004072. DOI: 10.1161/CIRCEP.116.004072.
 53. Marai I, Suleiman M, Blich M, et al. Clinical and electrophysiologic outcomes of patients undergoing percutaneous endocardial ablation of scar-related ventricular tachycardia: A single-center experience. *ISR Med Assoc J*. 2012;12:667–70.
 54. Volkmer M, Ouyang F, Deger F, et al. Substrate mapping vs. tachycardia mapping using CARTO in patients with coronary artery disease and ventricular tachycardia: Impact on outcome of catheter ablation. *Europace*. 2006;8:968–76. DOI: 10.1093/europace/eul109.
 55. Antz M, Berodt K, Bänsch D, et al. Catheter-ablation of ventricular tachycardia in patients with coronary artery disease: Influence of the endocardial substrate size on clinical outcome. *Clin Res Cardiol*. 2008;97:110–7. DOI: 10.1007/s00392-007-0596-7.
 56. Bogun F, Krishnan S, Siddiqui M, et al. Electrogram characteristics in postinfarction ventricular tachycardia: Effect of infarct age. *J Am Coll Cardiol*. 2005;46:667–74. DOI: 10.1016/j.jacc.2005.01.064.
 57. Bruckhorst CB, Delacretaz E, Soejima K, et al. Identification of the ventricular tachycardia isthmus after infarction by pace mapping. *Circulation*. 2004;110:652–9. DOI: 10.1161/01.CIR.0000138107.11518.AF.
 58. Deneke T, Grewe PH, Lawo T, et al. Substrate-modification using electroanatomical mapping in sinus rhythm to treat ventricular tachycardia in patients with ischemic cardiomyopathy. *Z Kardiol*. 2005;94:453–60. DOI: 10.1007/s00392-005-0240-3.
 59. Dinov B, Schönbauer R, Wojdyła-Hordynska A, et al. Long-term efficacy of single procedure remote magnetic catheter navigation for ablation of ischemic ventricular tachycardia: A retrospective study. *J Cardiovasc Electrophysiol*. 2012;23:499–505. DOI: 10.1111/j.1540-8167.2011.02243.x.
 60. Kettering K, Weig HJ, Reimold M, et al. Catheter ablation of ventricular tachycardia in patients with ischemic cardiomyopathy: Validation of voltage mapping criteria for substrate modification by myocardial viability assessment using FDG PET. *Clin Res Cardiol*. 2010;99:753–60. DOI: 10.1007/s00392-010-0182-2.
 61. Li YG, Grönefeld G, Israel C, et al. Stepwise approach to substrate modification of left ventricular scar. *Circ Arrhythm Electrophysiol*. 2015;8:76–86. DOI: 10.1161/CIRCEP.114.001827.
 62. Jamil-Copley S, Vergara P, Carubicchio C, et al. Application of ripple mapping to visualize slow conduction channels within the infarct-related left ventricular scar. *Circ Arrhythm Electrophysiol*. 2015;8:76–86. DOI: 10.1161/CIRCEP.114.001827.
 63. Verma A, Marrouche NF, Schweikert RA, et al. Relationship between successful ablation sites and the scar border zone defined by substrate mapping for ventricular tachycardia post-myocardial infarction. *J Cardiovasc Electrophysiol*. 2005;16:465–71. DOI: 10.1046/j.1540-8167.2005.40443.x.
 64. Miyamoto K, Noda T, Satomi K, et al. Larger low voltage zone in endocardial Unipolar map compared with that in epicardial bipolar map indicates difficulty in eliminating ventricular tachycardia by catheter ablation. *Heart Vessels*. 2016;31:1337–46. DOI: 10.1007/s00380-015-0732-7.
 65. Yamashina Y, Yagi T, Namekawa A, et al. Distribution of successful ablation sites of idiopathic right ventricular outflow tract tachycardia. *Pacing Clin Electrophysiol*. 2009;32:727–33. DOI: 10.1111/j.1540-8159.2009.02358.x.
 66. Parreira L, Cavaco D, Reis-Santos K, et al. Remote magnetic navigation for mapping and ablation of right and left ventricular outflow tract arrhythmias. *Rev Port Cardiol*. 2013;32:489–95. DOI: 10.1016/j.repc.2012.12.012.
 67. Yamashina Y, Yagi T, Namekawa A, et al. Clinical and electrophysiological difference between idiopathic right ventricular outflow tract arrhythmias and pulmonary artery arrhythmias. *J Cardiovasc Electrophysiol*. 2010;21:163–9. DOI: 10.1111/j.1540-8167.2009.01601.x.
 68. Tovia-Brodie O, Belhassen B, Glick A, et al. Use of new imaging CARTO(R) Segmentation Module software to facilitate ablation of ventricular arrhythmias. *J Cardiovasc Electrophysiol*. 2017;28:240–8. DOI: 10.1111/jce.13112.
 69. Verma A, Kilicaslan F, Schweikert RA, et al. Short- and long-term success of substrate-based mapping and ablation of ventricular tachycardia in arrhythmogenic right ventricular dysplasia. *Circulation*. 2005;111:3209–16. DOI: 10.1161/CIRCULATIONAHA.104.510503.
 70. Satomi K, Kurita T, Suyama K, et al. Catheter ablation of stable and unstable ventricular tachycardias in patients with arrhythmogenic right ventricular dysplasia. *J Cardiovasc Electrophysiol*. 2006;17:469–76. DOI: 10.1111/j.1540-8167.2006.00434.x.
 71. Martin CA, Takigawa M, Martin R, et al. "Use of novel electrogram "Lumipoint" algorithm to detect critical isthmus and abnormal potentials for ablation in ventricular tachycardia". *JACC Clin Electrophysiol*. 2019;5:470–9. DOI: 10.1016/j.jacep.2019.01.016.
 72. Nüchrich JM, Kaiser L, Akbulak RÖ, et al. Substrate characterization and catheter ablation in patients with scar-related ventricular tachycardia using ultra high-density 3-D mapping. *J Cardiovasc Electrophysiol*. 2017;28:1058–67. DOI: 10.1111/jce.13270.
 73. Sultan A, Bellmann B, Lüker J, et al. The use of a high-resolution mapping system may facilitate standard clinical practice in VE and VT ablation. *J Interv Card Electrophysiol*. 2019;55:287–95. DOI: 10.1007/s10840-019-00530-1.
 74. Viswanathan K, Mantziari L, Butcher C, et al. Evaluation of a novel high-resolution mapping system for catheter ablation of ventricular arrhythmias. *Heart Rhythm*. 2017;14:176–83. DOI: 10.1016/j.hrthm.2016.11.018.
 75. Lee W-C, Wu P-J, Fang H-Y, et al. Late fractionated potentials in catheter ablation for right ventricular outflow tract ventricular arrhythmias. *Pacing Clin Electrophysiol*. 2019;42:1115–24. DOI: 10.1111/pace.13748.
 76. Nayyar S, Wilson L, Ganesan AN, et al. High-density mapping of ventricular scar: A comparison of ventricular tachycardia (VT) supporting channels with channels that do not support VT. *Circ Arrhythm Electrophysiol*. 2014;7:90–8. DOI: 10.1161/CIRCEP.113.000882.
 77. Miyamoto K, Tsuchiya T, Narita S, et al. Radiofrequency catheter ablation of ventricular tachyarrhythmia under navigation using ensite array. *Circ J*. 2010;74:1322–31. DOI: 10.1253/circj.cj-09-1008.
 78. Nair M, Yaduvanshi A, Kataria V, Kumar M. Radiofrequency catheter ablation of ventricular tachycardia in arrhythmogenic right ventricular dysplasia/cardiomyopathy using non-contact electroanatomical mapping: Single-center experience with follow-up to median of 30 months. *J Interv Card Electrophysiol*. 2011;31:141–7. DOI: 10.1007/s10840-011-9556-2.
 79. Hocini M, Shah AJ, Neumann T, et al. Focal arrhythmia ablation determined by high-resolution noninvasive maps: Multicenter feasibility study. *J Cardiovasc Electrophysiol*. 2015;26:754–60. DOI: 10.1111/jce.12700.
 80. National Heart, Lung, and Blood Institute. Study quality assessment tools. 2019. Available at: www.nhlbi.nih.gov/health-topics/study-quality-assessment-tools (Date last accessed: 15 May 2021).
 81. National Institutes of health. Quality Assessment Tool for Observational Cohort and Cross-Sectional Studies. 2014. Available at: www.nhlbi.nih.gov/health-pro/guidelines/in-develop/cardiavascular-risk-reduction/tools/cohort (Date last accessed: 28 March 2024).
 82. Seitz J, Horvilleur J, Lacotte J, et al. Correlation between AF substrate ablation difficulty and left atrial fibrosis quantified by delayed-enhancement cardiac magnetic resonance. *Pacing Clin Electrophysiol*. 2011;34:1267–77. DOI: 10.1111/j.1540-8159.2011.03148.x.
 83. Shah AJ, Hocini M, Xhaet O, et al. Validation of novel 3-dimensional electroanatomical mapping of atrial tachycardias by invasive mapping and ablation: A multicenter study. *J Am Coll Cardiol*. 2013;62:889–97. DOI: 10.1016/j.jacc.2013.03.082.
 84. Gepstein L, Hayam G, Ben-Haim SA. A novel method for nonfluoroscopic catheter-based electroanatomical mapping of the heart. *Circulation*. 1997;95:1611–22. DOI: 10.1161/01.CIR.95.6.1611.
 85. Nakagawa H, Ikeda A, Sharma T, et al. Rapid high resolution electroanatomical mapping: Evaluation of a new system in a canine atrial linear lesion model. *Circ Arrhythm Electrophysiol*. 2012;5:417–24. DOI: 10.1161/CIRCEP.111.968602.
 86. Cen Z, Yang W, Xie Z, et al. Multi-electrode mapping of complex macroreentry atrial tachycardia. *J Electrocardiol*. 2020;60:27–32. DOI: 10.1016/j.jelectrocard.2019.11.039.
 87. Medical Advisory Secretariat. Advanced electrophysiologic mapping systems: An evidence-based analysis. *Ont Health Technol Assess Ser*. 2006;6:1–101.
 88. Liu X, Wang X, Gu J, et al. Electroanatomical systems to guided circumferential pulmonary veins ablation for atrial fibrillation: Initial experience from comparison between the EnSite/Navx and CARTO system. *Chin Med J (Engl)*. 2005;118:1156–60.
 89. Maury P, Champ-Rigot L, Rollin A, et al. Comparison between novel and standard high-density 3D electro-anatomical mapping systems for ablation of atrial tachycardia. *Heart Vessels*. 2019;34:801–8. DOI: 10.1007/s00380-018-1307-1.
 90. Rottner L, Metzner A, Ouyang F, et al. Direct comparison of point-by-point and rapid ultra-high-resolution electroanatomical mapping in patients scheduled for ablation of atrial fibrillation. *J Cardiovasc Electrophysiol*. 2017;28:289–97. DOI: 10.1111/jce.13160.
 91. Ikeda Y, Kato R, Mori H, et al. Impact of high-density mapping on outcome of the second ablation for atrial fibrillation. *J Interv Card Electrophysiol*. 2021;60:135–46. DOI: 10.1007/s10840-020-00716-y.
 92. Kaseno K, Hasegawa K, Miyazaki S, et al. Discrepancy between CARTO and Rhythmia maps for defining the left atrial low-voltage areas in atrial fibrillation ablation. *Heart Vessels*. 2021;36:1027–34. DOI: 10.1007/s00380-021-01773-7.
 93. Aliot EM, Stevenson WG, Almond-Garrote JM, et al. EHRA/HRS Expert Consensus on Catheter Ablation of Ventricular Arrhythmias: Developed in a partnership with the European Heart Rhythm Association (EHRA), a registered branch of the European Society of Cardiology (ESC), and the Heart Rhythm Society (HRS), in collaboration with the American College of Cardiology (ACC) and the American Heart Association (AHA). *Europace*. 2009;11:771–817. DOI: 10.1093/europace/eup098.